

3/2/16

Grow Portland

Re: Garden School

Dear Grow Portland

Enclosed are the results of analysis for samples received by the laboraory on $\frac{2}{26}$ 2016

The results related only to the samples included in this report.

The project was assigned a report number of 85769

If you have any questions concerning this report, please feel free to contact us.

Sincerely,

OV Ob ---

CY Chan QA Officer



Laboratory Report

Grow Portland

Project Name:

Project Location: Garden School

Project Number:

Date Sampled: 2/24/16

Date received:

2/26/16

EPA 7000

Report Date:

Report Number:

Analyte: Total Lead (Pb) quantitation by FLAA

85769

3/2/16

Field ID	Lab ID	Result	Detection Limit	
		mg/Kg(ppm)	mg/Kg(ppm)	
Ockley	J6175	46	17	
Chief Joseph	J6176	69	19	
Peninsula	J6177	34	18	

Quality Control Report: Lead (Pb)

	Measured Conc.	Expected Concentration	Recovery	Lower		
PB160302-1	(mg/L)	(mg/L)	(%)	Limit	Upper Limit	
ICV	5.2	5.0	104%	80%	120%	PASS
Prep Blank	0.0	0.0		0.5 mg/L		PASS
LCS	2.9	2.6	110%	70%	130%	PASS
QC Blank	0.0	0.0		0.5 mg/L		PASS
QC Check	5.3	5.0	105%	80%	120%	PASS
Calibration R^2 :	0.9999	(Lower Limit:	0.990)			PASS

Environmental Sciences, Inc.

CHAIN OF CUSTODY

Report Number

lumber

Relinquished by Site Project # Company / Samples: Project Name / 2415 SE 11th Ave. Portland Oregon 97214 JC 757 LAB ID March Temperature 18.1 (On loe? Yet / No DOY don 00c大 Portland on in Sula 5 CM 00 leu Field ID Affiliation Date Date FAX 0/24/W Phone Report Attention Purchase Order # J/14/16 Turnaround Time: Sampling Date 503-477-2333 Sampling Time abelleragrow portland, org Time Time Regular Collected By 3-5 Business Days Received by Received by Container Volume NW-TPH-Dx **NW-TPH-GX** W/Z Affiliation NW-TPH-HCID EPA 8021B (BTEX) EPA 8270 SIM (PAH) Phone(503) 231-9320 FAX(503) 231-9344 **EPA 8260B** Analysis Requested Comments Lego Lead Lead Time Pg. 3 of 3